



PHOTOMETRE 722

Analyseur de biochimie

- Wavelength range **330-1000nm**
- Spectral band 6nm
- Wavelength accuracy 2nm
- Stray light Less than 0.5%T@340nm
- Transmittance 0%to 125.0%T
- Absorbency 0A to 1.999A
- Concentration 0-1999C 0-1999F

- Power 220V 50Hz

